

177313-02-02-00

Electric TK-3e Proximity System Test Kit

Bently Nevada 177313-02-02-00 is an electric TK-3e proximity system test kit with metric micrometer and European power cord. It simulates shaft position and vibration for checking proximity transducer and monitor response.

Manufacturer	Bently Nevada
Product family	TK-3 Proximity System Test Kit
Part number	177313-02-02-00
Country of origin	USA
Source	TK3 Proximity System Test Kit Datasheet — 178087 Rev. G

Product role

- 1

Simulates shaft vibration for proximity-system verification
- 2

Provides a micrometer-controlled static target position
- 3

Generates an adjustable rotating wobble target
- 4

Supports controlled checks of probes, Proximitor Sensors and monitor channels

Document purpose

Technical review	Model identity and manufacturer-published values
Installation review	Mounting, connection and companion-hardware checks
Commissioning review	Controlled signal, diagnostic and functional response

APTER POWER SALES TEAM

WHATSAPP / PHONE
+86 180 3023 5313
EMAIL
sales13@apterpower.com
WEBSITE
www.apterpower.com



177313-02-02-00

1. Technical specifications

The following values are transcribed only from the manufacturer records named in this document.

Drive type	Electric
Micrometer units	Metric
Power cord	European
Approval option	None
Supported supply	95 to 125 Vac or 190 to 250 Vac
Supply frequency	50/60 Hz
Minimum supply current	1 A
Wobulator displacement	50 to 254 um peak-to-peak
Wobulator speed	0 to 5000 cpm +/-1000 cpm
Micrometer range	0 to 25.4 mm
Target material	AISI 4140 steel
Operating temperature	0 to +50 degC
Continuous operation maximum	+40 degC
Storage temperature	-18 to +65 degC
Relative humidity	95%
Enclosure rating	IP54 when closed
Universal probe mount	5 to 19 mm
Maximum operating altitude	2000 m

Configuration identity checks

Complete model	177313-02-02-00
Manufacturer part	177313-02-02-00
Installed family	TK-3 Proximity System Test Kit
Record authority	178087 Rev. G

2. System function and integration

Bently Nevada 177313-02-02-00 is an electric TK-3e proximity system test kit with metric micrometer and European power cord. It simulates shaft position and vibration for checking proximity transducer and monitor response.

Compatible installation elements

Component 1	compatible proximity probe
Component 2	Proximitor Sensor and extension cable
Component 3	machinery monitor channel
Component 4	portable measurement instrument

Engineered application matrix

Proximity probe calibration check	Simulates shaft vibration for proximity-system verification
Monitor channel vibration simulation	Provides a micrometer-controlled static target position
Static gap verification	Generates an adjustable rotating wobble target
Maintenance shop transducer testing	Supports controlled checks of probes, Proximitor Sensors and monitor channels

Related products on Apter Power

330106-05-30-05-02-00	3300 XL 8 mm reverse-mount probe	Open product page
330103-00-12-10-02-00	3300 XL 8 mm metric probe	Open product page
330103-00-16-10-02-05	3300 XL 8 mm approved probe	Open product page
330130-045-00-00	3300 XL extension cable	Open product page

3. Installation and commissioning controls

Secure the test probe in the universal mount, select the correct supply and establish static micrometer position before increasing wobulator speed. Compare electrical and monitor readings throughout the controlled test range.

- 1

Verify 177313-02-02-00, metric micrometer and European cord before use.
- 2

Inspect the enclosure, drive assembly and target surfaces before applying power.
- 3

Select only a supply range supported by the electric TK-3e.
- 4

Clamp the probe securely without damaging its case or cable.
- 5

Keep hands, leads and loose items clear of the rotating wobble plate.
- 6

Start at zero speed and increase only after confirming mechanical clearance.
- 7

Use the AISI 4140 targets when comparing published transducer response.
- 8

Return controls to a safe state and close the IP54 enclosure after testing.

Acceptance record

Identity	177313-02-02-00 recorded
Mounting / connection	compatible proximity probe
Companion hardware	Proximitior Sensor and extension cable
Functional duty	Proximity probe calibration check
As-left evidence	Connections, diagnostics and controlled response recorded

Return-to-service sequence

1	Confirm physical security and connector condition
2	Restore the documented channel or system configuration
3	Observe stable diagnostics before applying a controlled test
4	Record the final process response and release authorization

4. Engineer FAQ and source record

What maintenance evidence should be retained for 177313-02-02-00?	Retain the full 177313-02-02-00 identity, as-left mounting and connections, continuity or bias checks, diagnostics and the final functional-test result.
What should be verified before returning 177313-02-02-00 to service?	Confirm secure mounting, connector integrity, restored configuration, stable signal behavior and the intended response of the associated monitoring channel.
What must be matched before replacing 177313-02-02-00?	Match the complete 177313-02-02-00 marking, electric tk-3e proximity system test kit, installed mounting or connector execution and the approved equipment record before removal.
Which system components should be checked with 177313-02-02-00?	Confirm the installed compatible proximity probe, Proximitor Sensor and extension cable and the assigned monitoring or connection point are the documented combination for 177313-02-02-00.
How should wiring or cabling for 177313-02-02-00 be recorded?	Record conductor identity, shield termination, connector orientation, cable route and the as-found diagnostic state before disconnecting 177313-02-02-00.
What installation control is important for 177313-02-02-00?	Use the model-specific mounting arrangement, protect mating surfaces and maintain the documented cable support and mechanical clearance for 177313-02-02-00.
How is 177313-02-02-00 checked during commissioning?	Apply the approved controlled test for proximity probe calibration check, verify polarity or direction where applicable, and compare the measured response with the configured channel.
Can parameters from a similar model be used for 177313-02-02-00?	No. Retain only values and configuration details that apply directly to the complete 177313-02-02-00 ordering code and installed system.

Manufacturer source identifiers

Document title	TK3 Proximity System Test Kit Datasheet
Publication / revision	178087 Rev. G
Manufacturer	Bently Nevada / Baker Hughes

APTER POWER SALES TEAM

WHATSAPP / PHONE
+86 180 3023 5313

EMAIL
sales13@apterpower.com

WEBSITE
www.apterpower.com



177313-02-02-00